


**OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER,
DHENKANAL
E MAIL:- rbbdhenkanal@gmail.com
Phone No:-**

Letter No:-

Date:-

Tender Call Notice.

Sealed Tender is invited from authorized Distributor/Dealer/ Agency/ manufacturer For Supply & installation of Blood Donor Couch at **Blood Bank, DHH, Dhenkanal** , as per Following Attached Specification. The details are available in the district website:- www.dhenkanal.nic.in. The tender paper cost 500.00 (Non refundable) in shape of Demand Draft drawn from any nationalized bank in favor of, Blood Bank Officer I/C, Odisha Blood bank, Dhenkanal, Payble at Dhenkanal , should be enclosed with technical bid. The tender should submit on or before: -03-12-2018 till 4 Pm By Regd Post/ Speed Post/ Courier at Blood Bank, DHH, Dhenkanal Premises. The Tender Paper will be Opened on Same day by presence of Bidders/ or their Authorized Representative. Absence of Bidders or their representative wills no bar for opening of the Tender. The Tender Authority has the right to accept or reject part/full tender without assigning any reason thereof.


CDM & PHO

DHENKANAL

TERMS & CONDITIONS FOR SUPPLY OF BLOOD DONOR COUCH.

- The bidder's have to submit their tender in separate sealed cover (Cover A (technical Bid) and Cover B (Price Bid)) both the covers should be put in to a third cover C which must be super scribe as ' tender for supply and installation of Donor couch at Blood Bank DHH, Dhenkanal.

The cover A (Technical Bid) should contain

- 1 Check list Details of the documents with page no
- 2 Manufacturing/ Supplier/ Authorized Diller should submit an affidavit from notary public with satisfying non blacklisted for any state govt. or Central Govt./ PSU units.
- 3 E M D rupees 5000.00(Five thousand Only) (Refundable in safe of Bank Draft In Favor of Blood Bank Officer , Odisha Blood Bank Dhenkanal to be attached.
- 4 Tender paper cost of rupees 500.00(five Hundred non refundable) to be attached.
- 5 Copy of PAN Card.
- 6 Copy of IT Returns of last three years.
- 7 Copy of GST registration Certificate.
- 8 Authorization from the prinripal in case of Distributer.
- 9 Catalogue/ Technical Broachers of the products.

TECHNICAL SPECIFICATION

The Donor Couch Should be design to provide maximum comfort and safety to the patients with logically space bottom remort controller on easy to operate and allow quick adjustment best of individual preference. Have arm rest with optimally. The height adjustment is done vertically with help of reliable heavy duty leafing clumbs and it should be safety and comfort of the donor and convenience of the phlebotomy.

Height	620mm-820mm
Profile Length	2010mm
Seat width	680mm
Backrest Adjustment	0-54Degree
Leg rest Adjustment	0-60 Degree
Trendelenburg Position	15 Degree
Castrol whells	400mm with luch
Voltage	230volt
Accessories	2 accessories tray, Magazine rack and IV pole spigmano meter tray and blood collection monitor tray.

ANNUEXURE-I

FINANCIAL BID

Si No	Items	Sprcification	Quantity	Amount Including Tax
1	Height	620mm-820mm	2 No's	
2	Profile Length	2010mm		
3	Seat width	680mm		
4	Backrest Adjustment	0-54DegreeE		
5	Leg rest Adjustment	0-60 Degree		
6	Trendelenburg Position	15 Degree		
7	Castrol whells	400mm with luch		
8	Voltage	230volt		
9	Accessories	2 accessories tray, Magazine rack and IV pole spigmano meter tray and blood collection monitor tray.		

lms
20/11/18

 CDM & PHO

DHENKANAL