



ODISHA STATE AGRICULTURAL MARKETING BOARD

Plot No.1800(P), Baramunda, Post-Khandagiri, Near Indian Oil Petrol Pump Bhubaneswar
Ph. No.0674-2351311, Fax : 0674- 2351316, Email : osamboard99@yahoo.com, Website : osamboard.org

Letter No.3252(66)
Mkt-III-94/2017-18

Date.30.10.18

To

The Chairpersons,
All Regulated Marketing Committees in the State.

Sub Procurement of equipments for quality testing, weighment and other materials by RMCs for paddy procurement.

Ref Letter No.2997 dated.06.10.2018

Madam / Sir,

With reference to the subject cited above in the letter under reference, requirements of various equipments, materials for safe storage of goods and weighments were called from different RMCs, after the approval of the same by the District Level Purchase Committee under the Collector. Information in this regard has been received only from some RMCs, and other RMCs are yet to submit the information.

This information is being obtained to consider whether centralized procurement at the level of OSAM Board may be done, on which further decision has not yet been taken. However, the specifications for the equipments / materials have already been finalized and the same are enclosed for your reference.

In the mean time, procurement activities are going to start from the coming month and some RMCs have asked for clarification in this regard. Almost all the RMCs have reported some stock of such equipments and materials already available to take care of initial requirements for paddy procurement operations. However, in order to avoid any dislocation, the RMCs are instructed that they may undertake procurement of equipments and materials meeting the specifications enclosed to take care of the immediate requirements, if any. However, instructions given with regard to stock entries, due verification and other aspects to ensure procurement of good quality equipments / materials, as issued in this office letter No.2997 dated 06.10.2018 must be followed scrupulously, failing which appropriate action may be initiated against the Secretary of RMC concerned.

Yours faithfully,

Encl: 1 + 0 3 pages

[Handwritten Signature]
20/10/18

Memo No. 3253 /Date 30.10.18

Copy forwarded to all Collectors in the State for information and necessary action.

[Signature]
29/10/18

Director, Agricultural Marketing-Cum-
Member Secretary, OSAM Board

Memo No. 3254 /Date 30.10.18

Copy forwarded to Private Secretary to Hon'ble Chairman, OSAM Board for favour of kind information of Hon'ble Chairman.

[Signature]

Director, Agricultural Marketing-Cum-
Member Secretary, OSAM Board.

Memo No. 3255 /Date 30.10.18

Copy forwarded to Private Secretary to Commissioner-cum-Secretary to Government, Cooperation Department for favour of kind information of Commissioner-Cum-Secretary.

[Signature]

Director, Agricultural Marketing-Cum-
Member Secretary, OSAM Board

TECHNICAL SPECIFICATION OF MINI SEED GRADER

Sl. No.	Specifications
1	Laboratory Model
2	Commodity : Paddy, Pulses
3	Operating Voltage : 220 Volt, 50 Hz A.C
4	Screen area : 29 x 24 cm 23pprox.
5	Screen provided : Set of 10 Sieves
6	Sieve sizes (mm): 3.5 OH, 2.75 OH, 2.0 OH, 1.85 OH, 1.75 OH, 1.5 OH, 7.0 RH, 5.0 RH, 1.6 RH and 1.2 RH (OH-Oblong Hole and RH - Round Hole)
7	No. of working screen : Two
8	Motor: ½ HP Single Phase (ISI Standard)
9	Operation : Manual, Power
10	Adjustable feed control, Air speed regulator, Stainless Steel Sieve
11	Made of teak wood and ply.

TECHNICAL SPECIFICATION OF SET OF SIEVES

Sl. No.	Specifications
1	Set of Sieves (4 nos. with collecting pan and lid made up of brass)
2	Sieve Size - 4 mm - 1 no. 3.35 mm - 1 no. 1.7 mm - 1 no. 1 mm - 1 no. Dust collector - 1 no
3	Solid Bottom Pan - 1 no.
4	Lid - 1 no.

TECHNICAL SPECIFICATION OF DIGITAL MOISTURE METER

Sl. No.	Specifications
1	Measuring Range: 3.5% to 40%
2	Supply Voltage : 9V DC
3	Temperature Compensation : Automatic calibration & error correction facility
4	Display : Digital LCD / LED
5	Error condition display for over Range or under range : $\pm 0.2\%$
6	Commodity : Cereals, Pulses, Oil Seeds
7	Operating Temperature : Ambient to 50° C

TECHNICAL SPECIFICATION OF ANALYSIS BOX / KIT

Sl. No.	Specification
1.	Analysis Box / Kit consists of Class 'B' Mechanical Weighing Balance up to 100 gm. Weight box consist of set of verified weights up to 100 gms. Enamelled Plate-9", Parkhi-9", Sieve-3.30 mesh, Magnifying Glass-1.5" (Magnification of 10X), Vernier / Micrometer, Palm Paddy De-husker, Scoop (With handle, 105 mm x 1000 mm x 25 mm, Measuring Cylinder-20 cc., Built-in Purity Work Board for testing physical purity of seeds, Petri Dish & Forceps all packed in a compact wooden / leather brief case, as small, as light and as compact as possible, but as accurate and as reliable.

TECHNICAL SPECIFICATION OF MICRO METRE

Sl. No.	Specifications
1	Analog measuring device for measurement of length / breadth of seed. The dial used is as per IS specification.
2	Made of stainless steel
3	Range 0 – 25 mm
4	Resolution : 0.01 mm
5	Accuracy : 0.005 mm

TECHNICAL SPECIFICATION OF SEED SAMPLE DIVIDER (BOERNER TYPE)

Sl. No.	Specifications
1.	Meeting BIS specification
2	No. of Pockets : 38
3	Hopper Capacity : 1 Kg (minimum)
4	Feature : provides accurate splits of the grain Samples into two equal halves with an accuracy of $\pm 5\%$ on a 1Kg. sample.
5.	Two collecting Pans
6	Copper and Brass Body
7	Made without joint (in material flowing component)

TARPAULIN & HDPE POLYTHENE

To be procured from valid rate contract of Director, Export Promotion & Marketing (DEP&M), Odisha, Bhubaneswar as per their approved specification and rate contract of Tarpaulin and HDPE Polythene.